Non-Integer Root Transformations for Preprocessing Nano-Electrospray Ionization High
Resolution Mass Spectra for the Classification of Cannabis

A thesis presented to
the faculty of
the College of Arts and Sciences of Ohio University

In partial fulfillment
of the requirements for the degree
Master of Science

Yue Tang

August 2018

© 2018 Yue Tang. All Rights Reserved.

This thesis titled
Non-Integer Root Transformations for Preprocessing Nano-Electrospray Ionization High
Resolution Mass Spectra for the Classification of Cannabis

by
Yue Tang

has been approved for
the Department of Chemistry and Biochemistry
and the College of Arts and Sciences by

Peter B. Harrington
Professor of Chemistry and Biochemistry

Joseph Shields
Interim Dean of Arts and Sciences

ABSTRACT

Tang, Yue., M.S., August 2018, Chemistry

Non-Integer Root Transformations for Preprocessing Nano-Electrospray Ionization High Resolution Mass Spectra for the Classification of Cannabis

Director of Thesis: Peter B. Harrington

Typically, for measurements with high dynamic range, the range is reduced by using the square root transform. By using non-integer roots coupled with systematic experimental design, improvements to the measurements may be obtained. The effect of using non-integer root transformation was evaluated using high-resolution mass spectrometry (HRMS) combined with nano-electrospray ionization (Nano-ESI) to differentiate 23-samples of *Cannabis*. The mass spectra were evaluated and classified using different mass resolving powers and non-integer root transformations. Classification was achieved by super partial least squares discriminant analysis (sPLS-DA), support vector machine (SVM), and SVM classification tree type entropy (SVMTreeH). The 2.5 root transformation gave the best overall performance at different resolving powers for chemical profiling from a multilevel (4×5) factorial experimental design. Response surface modeling using a cubic polynomial model of the sPLS-DA prediction accuracies yielded optima at 0.00407 for resolving power and 2.51 for the root transformation. Root transformation is an important spectral preprocessing tool for decreasing the dynamic range so that the relative variance of smaller but more important features may be inflated.

ACKNOWLEDGMENTS

Chemistry Mapping, Inc. and OHIO Center for Intelligent Chemical Instrumentation partially support this research. Chemistry Mapping, Inc. is acknowledged for supplying the twenty-three *Cannabis* extracts.

TABLE OF CONTENTS

	Page
Abstract	3
Acknowledgments.....	4
List of Tables	6
List of Figures	7
Introduction.....	8
Theory	错误!未定义书签。 2
Root transform	12
Super partial least squares discriminant analysis	12
Support vector machine	12
Support vector machine classification trees type entropy.....	13
Cubic polynomial model.....	13
Bootstrapped Latin partitions.....	14
Principal component analysis	14
Experimental Section	16
Discussion of Results.....	19
Conclusions.....	26
Literature Cited	27

LIST OF TABLES

	Page
Table 1: Commercial names of the <i>Cannabis</i> products extracted in CDCl_3	16
Table 2: The sPLS-DA Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions.....	23
Table 3: The SVM Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions	23
Table 4: The SVMTreeH Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions	23
Table 5: Process of Chasing the Optima of the Cubic Response Surface Model of the Average sPLS-DA Percent Accuracies	25

LIST OF FIGURES

	Page
Figure 1: Chromatogram averaged and normalized mass spectra with 10^{-2} amu bin size or resolving power of the <i>Cannabis</i> extracts.....	19
Figure 2: PCA scores for the chromatogram averaged mass spectra from the 10^{-2} linear binning of the 23 <i>Cannabis</i> extracts and their 5 replicates	21
Figure 3: SVMTreeH classification tree of the 23 <i>Cannabis</i> samples with 10^{-2} linear binning with 2.5 root transformation.....	22
Figure 4: Estimated classification accuracy from fitted cubic polynomial of the sPLS-DA classification accuracies	24

INTRODUCTION

In some states of the United States, two species of *Cannabis*, *Cannabis sativa* and *Cannabis indica* have been legalized for medical and recreational use¹. Medicinal *Cannabis* can be applied to treat a wide range of illnesses including loss of appetite, inflammation, pain, seizures, substance abuse disorders, mental disorders, and diseases that affect the immune system. There are over 400 different chemical compounds that have been identified in *Cannabis*, which includes mono- and sesquiterpenes, sugars, hydrocarbons, steroids, flavonoids, nitrogenous compounds, and amino acids². Various cannabinoids have received much attention for their pharmacological activities, such as tetrahydrocannabinol (THC), tetrahydrocannabinolic acid (THCa), cannabidiol (CBD), cannabinol (CBN), and cannabichromene (CBC)³. Although, many of the other components of *Cannabis* may be important cofactors that have medicinal affects and can mitigate side-effects of the purified active components. The medicinal benefit of a whole botanical medication is referred to as the Entourage Effect.

There are more than 700 different cultivated varieties of *Cannabis* that have been cataloged and this number is growing⁴. *Cannabis* is a polymorphic species that can hybridize to produce fertile offspring⁵. As the medicinal usage of *Cannabis* increases, it is necessary to clearly distinguish *Cannabis* by its expected therapeutic effects. One avenue is to characterize *Cannabis* products by their chemical profiles or chemotyping. To characterize *Cannabis* plant materials more accurately, many tools and methods have been applied which include nuclear magnetic resonance (NMR) spectroscopy profiling and high-resolution mass spectrometry (HRMS)⁶⁻⁸. The HRMS instruments such as the

Fourier-transform ion cyclotron resonance (FT-ICR) and Orbitrap provide selective and sensitive detection of the trace components of *Cannabis* with mass accuracies of a few parts per million (ppm). The Orbitrap detects trapped ions and measures the frequency of ion movement to image the mass to charge ratio (m/z) through the Fourier transform. The orbitrap also can identify ions using collision-induced dissociation (CID) and the secondary ion spectra⁹.

Nano-electrospray ionization (Nano-ESI) is one of the most efficient methods to detect liquid samples by mass spectrometry¹⁰. High voltage is applied to a liquid sample as it exits a narrow capillary to create microdroplets. The high voltage creates charges that reside on the droplet surface. As the droplet is heated, the evaporation decreases the size of droplets and Coulombic repulsion of the surface charges cause the droplets to break apart into smaller droplets with greater surface areas which facilitates the further evaporation of the solvent.¹¹

Compared to other ESI methods, Nano-ESI requires an extremely small sample size of microliters¹². One microliter of sample is injected into a glass capillary that has a tip diameter of 1 μm and is sprayed from the tip by applying high voltage (kV) to the solution¹³. The flow rate is a few nL/min. Compared with ESI, Nano-ESI reduces the interference effects of the analyte with salts, therefore the detection sensitivity is improved¹⁴. Those advantages make Nano-ESI a robust method for the classification of *Cannabis* and other botanical materials, especially because of the low sample consumption and high detection sensitivity.

For multivariate analysis, it is assumed that noise and ion abundance is uniformly distributed with respect to m/z . When this assumption is not satisfied, this condition is referred to as heteroscedastic, i.e., nonuniform distribution of the noise.

Transformation can be applied to mitigate the effect of heteroscedasticity and improve uniformity of the distribution of ion abundances¹⁵. Transformations are nonlinear conversion of the spectral intensities which include logarithmic and roots¹⁶. For the logarithmic transformation, mass intensities that are very small (i.e., close to zero) tend to over inflate, so that it does not work well for mass spectra. The square root transformation is often used as a heuristic to improve the ion abundance distribution. The integer root transformations, square root, cubic root, and quartic root, have been reported for the *Cannabis* samples measured on the Orbitrap instrument⁸. Our paper demonstrates the use of non-integer root transformations and the optimization of the transform with respect to classification accuracy using response surface modeling.

To reduce the experimental variations like sample size and concentration, normalization is required to pretreat the spectra¹⁷⁻¹⁸. This mass spectral preprocessing step seeks to accentuates the differences among the mass spectra from different samples and minimizes the differences among the replicate samples.

For our work, twenty-three *Cannabis* samples have been evaluation with Nano-ESI. For each sample, five replicates were collected, each on a separate day, using a random block design. The blocking variable between each replicate was one day.

The *Cannabis* spectra were mapped into the same Cartesian coordinate system for pattern recognition using a resolving power based binning method⁸. Resolving power

equals bin size m/z . The *Cannabis* spectra, through linear binning with mass resolving powers of 1, 0.1, 0.01, and 0.001, were mapped into a matrix representation so that each of the m rows of the matrix was a spectrum and each of the n columns of the matrix was a mass measurement at a specific mass-to-charge ratio (m/z).

The dynamic range of the spectra were decreased by applying the non-integer root transform to the MS data matrix using a 4×5 factorial experimental design¹⁹. The bootstrapped Latin partitions method was applied with 10 bootstraps and 2-Latin partitions to each design point to furnish a generalized and unbiased measure of classification accuracy. Each spectrum (i.e., row) of the data matrix was normalized to unit vector length. The effects of binning (i.e., mass resolution) and root transform were assessed with three different classifiers²⁰⁻²²; super partial least squares discriminant analysis (sPLS-DA), support vector machine (SVM), and support vector machine classification trees type entropy (SVMTreeH). The average classification rate from the 10 bootstraps was obtained for differentiating the 115 mass spectra into one of the 23 classes. The average classification accuracies for each design point were used for response surface modeling using a cubic polynomial¹⁹. Further experimental design points were added sequentially to chase the optima until the response surface model converged to a consistent set of optimal design points.

THEORY

1. Root transform

MS instruments generally have a large dynamic range. The peaks at higher masses have lower ion abundances but typically contain more characteristic or selective information⁸. Using the root transformation, the influence of the lower mass peaks with higher intensity will be attenuated and the higher mass peaks will more influential in building the classification model. For our study, the untransformed, 1.5 root, square root, 2.5 root, and cubic root transformations were compared and evaluated at different mass resolving powers.

2. Super partial least squares discriminant analysis

Super partial least squares discriminant analysis (sPLS-DA) is a supervised multivariate classification method²³. In the sPLS-DA algorithm, two equally sized partitions are randomly selected for the calibration spectra and their class labels²⁰. Each random split of the spectra is used to measure the prediction error and these prediction errors are pooled to furnish a comprehensive figure of merit. Because the data are split randomly, this process is repeated ten times and the pooled prediction error is stored with respect to the range of latent variables. The averaging across the bootstraps tends to smooth the curve of the prediction error with respect to latent variable number. The minimum average pooled prediction error across the ten bootstraps is used to define the number of latent variables. This number of latent variables is then used to construct the sPLS model from the entire set of calibration spectra.

3. Support vector machine

The support vector machine (SVM) is a learning algorithm that finds a decision plane that maximizes the separation among the closest points in the mass spectral data space (i.e., space defined by the mass spectra or rows of the data matrix) of spectra that belong to separate classes²⁴. Because the SVM is a binary classifier and can separate only two classes, the one against all strategy was used²¹. The one against strategy creates a separate SVM model for each class so that the target class is positive and all the other classes are negative²⁵. The set of SVM models are applied to an unknown spectrum and the model that yields the largest response will designate the predicted class (e.g., winner take all).

4. Support vector machine classification trees type entropy

Support vector machine classification trees type entropy (SVMTreeH) is a classification tree that comprises SVMs at each branch²². There are several advantages of these trees. First, they allow SVMs to be used for more than two classes. The encoding at each branch of the tree is optimized using the entropy of classification which allows trees to be constructed by having classes with largest number of objects separated first and may accommodate overlapping clusters of spectra in the dataspace. Lastly, because the objects are encoded optimally, there is no need for the SVM cost factor which greatly speeds up the SVM calculation.

5. Cubic polynomial model

The cubic polynomial model is a type of response surface model (RSM). RSM explores the relationships between explanatory variables and the response variables²⁶.

$$y = b_0 + b_1x_1 + b_2x_2 + b_3x_1^2 + b_4x_2^2 + b_5x_1x_2 + b_6x_1^3 + b_7x_2^3 + b_8x_1x_2^2 + b_9x_1^2x_2$$

For the cubic polynomial model, the coefficients $\mathbf{b}$ are obtained by regression of the response variable y onto the polynomial expanded design points x_1 and x_2 . Through the analysis of the regression equation, the optimal process parameters can be determined. In the provided graphs of the RSM, the optimization region is displayed. The resolving powers were converted to p-notation (i.e., the negative of the base-ten logarithm) so that they would be on the same scale at the root values.

6. Bootstrapped Latin partitions

Bootstrapped Latin partitions is a validation method for unbiased evaluation of chemometric models^{27 28}. For this method, the mass spectra are randomly split into model building and validation sets. The splitting procedure is constrained to maintain the same distribution of classes, because if the class distribution changes so will the model. Each split of the data is used once for validation and the other times assembled into a model-building set of data, so if the data are split into 25% proportions each split will be used once for validation and three times for model building. The results from the four validation sets are pooled so comprehensive statistics may be obtained. Because the splitting process is random, it may be repeated which is the bootstrap, i.e., the random selection of the pooled validation results. Averages and standard deviations may be calculated from the bootstrapped results.

7. Principal component analysis

Principal component analysis (PCA) is a statistical procedure for characterizing covariance to apply to MS study²⁹. The principal components are eigenvectors that are orthogonal and characterize the covariance of the mass spectra. Projection of the mass

spectra onto the eigenvectors yields scores that characterize the relationships among the mass spectra.

EXPERIMENTAL SECTION

Table 1. Commercial names of the <i>Cannabis</i> products extracted in CDCl ₃ .		
Type of <i>Cannabis</i>	Species (<i>indica:sativa</i>)	Label
Critical Mass	80 : 20	A
Blue Dream	60 : 40	B
Violator Kush	80 : 20	C
Sour Diesel	90 : 10	D
LSD	55 : 45	E
Lavender	60 : 40	F
Purple Durango OG	80 : 20	G
Headband	60 : 40	H
Gorilla Glue #4	50 : 50	I
Flo	60 : 40	J
Hindu Kush	100 : 0	K
Grape Kush	60 : 40	L
Critical Kush	90 : 10	M
Church	50 : 50	N
Ingrid	75 : 25	O
Lemon Skunk	60 : 40	P
GC Sour Diesel	0 : 100	Q
Solace Meds Sour Diesel	0 : 100	R
Good Lavender	60 : 40	S
GC Blue Widow	50 : 50	T
GC Grizzly Purple Kush	100 : 0	U
THCA + Extract Sour Diesel Terpene blend	0 : 100	V
THCA + True Terpene Sour Diesel Terpene Blend	0 : 100	W

The twenty-three *Cannabis* CDCl₃ extracts were provided by Chemistry Mapping, Inc. (Golden, CO) in deuterated chloroform so they may be subjected to NMR analysis. Aliquots of 200 μ L each of the *Cannabis* samples were diluted with 1.6 mL of HPLC grade methanol from Fisher Scientific (Houston, TX). Then 20 μ L of formic acid from Fisher Scientific (Houston, TX) was added as protonating reagent. *Cannabis* samples could contain more than 400 different chemical compounds. Each prepared sample was

filtered with a 0.02- μm Anotop 10 syringe filter from Whatman to remove particulates that may clog the Nano-ESI capillary.

The Thermo Scientific Q Exactive Plus^{TM} mass spectrometer (Bremen, Germany) was equipped with a Nano-ESI source and operated in positive ionization mode. A 500 μL gastight syringe from Hamilton was connected with a capillary from Polymicro Technologies to the ionization source. The glass capillary had a 100.2 μm diameter. The capillary was pulled with a laser micropipette puller, Model P-2000 (Sutter instruments Co., Novato, CA) so that one end of the capillary had a 1 μm diameter. The Nano-ESI measurements were obtained at 3 kV and 2 $\mu\text{L}/\text{min}$ flowrates for 120 s per sample. To obtain the *Cannabis* spectra, the scan range was m/z 150.0—2250.0, and the resolution was 280,000 at a scan rate of 0.9 Hz. Random block designs were applied with time as the blocking variable. Each block of replicates was collected on a different day.

The Thermo Q Exactive Plus^{TM} version 2.5 software was used to acquire the spectra. The spectra were saved as RAW data files and converted to the common document format (CDF) by Thermo Scientific Xcalibur^{TM} version 3.0.63 software. The CDF data was imported into MATLAB (MathWorks, Natick, MA) and stored in the mat file format.

Data matrices of spectra were constructed through linear binning from 1 to 0.001 resolving power. All the computations were performed using MATLAB R2018a (MathWorks Inc., Natick, MA) on personal computers with a Core i7 940 2.93 GHz with 8.0 GB of memory or a Core i7-3930K processor with 6 cores and 12 logical processors with 64 GB of memory (Intel, Santa Clara, CA). The former computer implemented

Windows 7 and the latter MS Windows 8 Enterprise 64-bit operating systems (Microsoft Corp., Redmond, WA). The computation time for the MS data analysis with different classifiers depended mostly on the resolving power which influence the number of variables n of the dataset. For the resolving power of 1 and a 1 m/z bin width, 2,103 columns of the data matrix were produced, and the MATLAB analysis time was 25 min. As the resolving power decreased to 0.001 and a 0.001 bin width, the number of columns for the data matrix increased to 1,637,449, so the analysis time increased to two hours.

DISCUSSION OF RESULTS

Fig. 1. Chromatogram averaged and normalized mass spectra with 10^{-2} amu bin size or resolving power of the *Cannabis* extracts. A) Untransformed, B) 1.5 root, C) 2 root, D) 2.5 root, E) 3 root.

The 115 *Cannabis* spectra were treated using five different root transformations. Fig. 1 is the chromatogram averaged and normalized mass spectra with the $10^{2.5}$ linear binning. For the linear binning spectra, in Fig. 1A, the untransformed mass spectra have high intensity peaks in the low m/z range, from m/z 100 to 500, but low intensity peaks in the higher m/z range.

Root transformations were applied to attenuate peak amplitudes in the lower m/z range. However, the root transform also will decrease the signal-to-noise ratios of the peaks in the spectra. The root transformations were from 1.5 to 3. In Fig 1c, after the square root transformation, more peaks are apparent in the range from m/z 500 to m/z 1000. In Fig 1d, after the 2.5 root transformation, peaks in the range of m/z 2000 are apparent. In Fig 1e, there are few additional peaks that appear compared to Fig 1d. Above the cubic root transformation the dynamic range was reduced too much, and noise began to dominate the spectra.

Fig. 2. PCA scores for the chromatogram averaged mass spectra from the 10^{-2} linear binning of the 23 *Cannabis* extracts and their 5 replicates. A) Untransformed, B) 1.5 root transformed, C) 2 root transformed, D) 2.5 root transformed, E) 3 root transformed.

Fig. 2 gives the PCA scores for the mass spectra using 10^{-2} resolving power with respect to the different transforms. In Fig. 2A, the sample scores for the untransformed spectra are not separated. After transformation, the separation of the sample scores improves. As the root increases, the scores begin to separate into well-defined classes that correspond to the samples, especially for the classes R, T, U, V, and W, which are at the lower-left quadrants of these figures.

Fig. 3 is the SVMTreeH classification tree of the 23 *Cannabis* samples with 10^{-2} linear binning with the 2.5 root transformation. The classification tree also indicated that all the sample spectra could be linearly separated with the SVMs at each branch of the

tree, because 22 SVM discriminants were required to separate the 23 samples. The distribution of classification tree leaves corresponded with the distribution of the sample score clusters in Fig 2.

Table 2. The sPLS-DA Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions.					
Root/Resolving Power	1	1.5	2	2.5	3
1	85±4	99.4±0.3	99.8±0.3	100±0	100±0
0.1	86±4	99.8±0.3	99.7±0.3	99.9±0.20	100±0
0.01	78±4	99.9±0.2	100±0	100±0	99.8±0.3
0.001	66±2	99.4±0.7	100±0	100±0	99.8±0.3

Table 3. The SVM Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions.					
Root/Resolving Power	1	1.5	2	2.5	3
1	94±1	98.6±0.8	99.6±0.4	99.7±0.5	99.6±0.5
0.1	96±1	99.3±0.9	99.8±0.4	99.8±0.4	99.5±0.7
0.01	96±1	99.2±0.5	99.9±0.2	99.9±0.2	99.6±0.5
0.001	85±2	98±1	99.7±0.5	99.9±0.2	99.4±0.7

Table 4. The SVMTreeH Classifier Average Percent Accuracies for the 23 <i>Cannabis</i> Classes of Spectra Using 10 Bootstraps and 2-Latin Partitions.					
Root/Resolving Power	1	1.5	2	2.5	3
1	87±2	97±1	99.2±0.6	99.7±0.4	99.9±0.2
0.1	90±2	99±1	99.6±0.5	99.4±0.7	99.3±0.7
0.01	89±3	98.7±0.6	99.8±0.4	99.8±0.4	99.0±1.0
0.001	78±4	97±2	99.7±0.4	99.8±0.3	97.7±0.8

Tables 2 to 4 are the classification accuracies from four different classifiers; PLS-DA, SVM, and SVMTreeH. In all three tables, 2.5 root and 0.01 resolving power gave the best classification accuracies within the 95% confidence intervals of the average

classification accuracies. These results indicate that the 2.5 root transformation had the best performance especially for sPLS-DA and the SVM.

Table 5. Process of Chasing the Optima of the Cubic Response Surface Model of the Average sPLS-DA Percent Accuracies.			
Root/Resolving Power	sPLS-DA	SVM	SVMTreeH
2.5/0.01	100±0	99.9±0.2	99.8±0.4
2.39/0.00457	100±0	99.8±0.3	99.8±0.3
2.85/0.00355	99.8±0.4	99.5±0.5	99.6±0.4
2.66/0.00389	100±0	99.9±0.2	99.8±0.3
2.71/0.00380	100±0	99.7±0.3	99.6±0.5
2.76/0.00372	100±0	100±0	99.8±0.3
2.51/0.00407	100±0	100±0	99.9±0.2

In Fig. 4, A cubic polynomial model was fit to the sPLS-DA results over a range of resolving powers and root transformations. After optimization of the response surface model, using the cubic regression equation, the optimal conditions of 0.00457 for resolving power and 2.39 for the root transformation were obtained. These optima were added as new experimental design points, and a new fit to the polynomial was calculated. Repeating this process until the optima converge is referred to as chasing the optima.

In Table 5, the chasing process was repeated for 5 times until the optima converged. In this table of the cubic polynomial model of sPLS-DA *Cannabis* results, the SVM and SVMTreeH results were also calculated using the same optima. So, the optima would be general and benefit the other two diverse classifiers that were unused during the optimization process. From the first and last rows of Table 5, an improvement in performance can be seen from the chasing the optima process for all three classifiers.

CONCLUSIONS

Chemical profiling by Nano-ESI combined with high-resolution mass spectrometry was applied to twenty-three *Cannabis* extracts and evaluated with five different root transformations, four different resolving powers and three different classifiers. The 2.5 root transformation, non-integer root, had the best performance in all the classifier results. The linear binning 0.01 and 0.001 resolving powers provides the best classification results. The best resolving powers and root transformations were estimated to be 0.00407 resolving power and 2.51 root transformation as obtained from the polynomial model fit to the sPLS-DA average classification rates. This study demonstrates that by using non-integer roots better preprocessing of mass spectra can be obtained. Furthermore, RSM is an efficient approach to optimize mass spectral preprocessing especially for optimizing the resolving power (i.e., bin width) of high resolution mass spectra and the dynamic range of the peak intensities. Lastly, it is demonstrated that too small a bin size can deleteriously affect the classification performance.

LITERATURE CITED

1. Hazekamp, A.; Tejkalová, K.; Papadimitriou, S., Cannabis: from cultivar to chemovar II—a metabolomics approach to Cannabis classification. *Cannabis and Cannabinoid Research* **2016**, *1* (1), 202-215.
2. ElSohly, M. A.; Slade, D., Chemical constituents of marijuana: The complex mixture of natural cannabinoids. *Life Sciences* **2005**, *78* (5), 539-548.
3. De Petrocellis, L.; Ligresti, A.; Moriello, A. S.; Allarà, M.; Bisogno, T.; Petrosino, S.; Stott, C. G.; Di Marzo, V., Effects of cannabinoids and cannabinoid - enriched Cannabis extracts on TRP channels and endocannabinoid metabolic enzymes. *British journal of pharmacology* **2011**, *163* (7), 1479-1494.
4. Hazekamp, A.; Fishedick, J., Cannabis - from cultivar to chemovar. *Drug testing and analysis* **2012**, *4* (7-8), 660-667.
5. Piomelli, D.; Russo, E. B., The Cannabis sativa versus Cannabis indica debate: an interview with Ethan Russo, MD. *Cannabis and cannabinoid research* **2016**, *1* (1), 44-46.
6. Harrington, P. d. B.; Wang, X., Spectral Representation of Proton NMR Spectroscopy for the Pattern Recognition of Complex Materials. *Journal of Analysis and Testing* **2017**, *1* (2), 10.
7. Joo, J. W. J.; Na, S.; Baek, J.-H.; Lee, C.; Paek, E., Target-Decoy with Mass Binning: a simple and effective validation method for shotgun proteomics using high resolution mass spectrometry. *Journal of proteome research* **2009**, *9* (2), 1150-1156.
8. Wang, X.; Harrington, P. d. B.; Baugh, S. F., Effect of preprocessing high-resolution mass spectra on the pattern recognition of Cannabis, hemp, and liquor. *Talanta* **2018**, *180*, 229-238.
9. Ma, Y.; Li, Q.; Van den Heuvel, H.; Claeys, M., Characterization of flavone and flavonol aglycones by collision - induced dissociation tandem mass spectrometry. *Rapid communications in mass spectrometry* **1997**, *11* (12), 1357-1364.
10. El-Faramawy, A.; Siu, K. M.; Thomson, B. A., Efficiency of nano-electrospray ionization. *Journal of the American Society for Mass Spectrometry* **2005**, *16* (10), 1702-1707.
11. Cole, R. B., Some tenets pertaining to electrospray ionization mass spectrometry. *Journal of Mass Spectrometry* **2000**, *35* (7), 763-772.
12. Schmidt, A.; Karas, M.; Dülcks, T., Effect of different solution flow rates on analyte ion signals in nano-ESI MS, or: when does ESI turn into nano-ESI? *Journal of the*

American Society for Mass Spectrometry **2003**, 14 (5), 492-500.

13. Karas, M.; Bahr, U.; Dülcks, T., Nano-electrospray ionization mass spectrometry: addressing analytical problems beyond routine. *Fresenius' journal of analytical chemistry* **2000**, 366 (6-7), 669-676.
14. Juraschek, R.; Dülcks, T.; Karas, M., Nanoelectrospray—more than just a minimized-flow electrospray ionization source. *Journal of the American Society for Mass Spectrometry* **1999**, 10 (4), 300-308.
15. Kvalheim, O. M.; Brakstad, F.; Liang, Y., Preprocessing of analytical profiles in the presence of homoscedastic or heteroscedastic noise. *Analytical Chemistry* **1994**, 66 (1), 43-51.
16. van den Berg, R. A.; Hoefsloot, H. C.; Westerhuis, J. A.; Smilde, A. K.; van der Werf, M. J., Centering, scaling, and transformations: improving the biological information content of metabolomics data. *BMC genomics* **2006**, 7 (1), 142.
17. Castillo, S.; Gopalacharyulu, P.; Yetukuri, L.; Orešič, M., Algorithms and tools for the preprocessing of LC–MS metabolomics data. *Chemometrics and Intelligent Laboratory Systems* **2011**, 108 (1), 23-32.
18. Yi, L.; Dong, N.; Yun, Y.; Deng, B.; Ren, D.; Liu, S.; Liang, Y., Chemometric methods in data processing of mass spectrometry-based metabolomics: a review. *Analytica chimica acta* **2016**, 914, 17-34.
19. Morgan, E.; Burton, K. W.; Church, P. A., Practical exploratory experimental designs. *Chemometrics and intelligent laboratory systems* **1989**, 5 (4), 283-302.
20. de B. Harrington, P.; Kister, J.; Artaud, J.; Dupuy, N., Automated principal component-based orthogonal signal correction applied to fused near infrared– mid-infrared spectra of French olive oils. *Analytical chemistry* **2009**, 81 (17), 7160-7169.
21. de Boves Harrington, P., Support vector machine classification trees. *Analytical chemistry* **2015**, 87 (21), 11065-11071.
22. de Boves Harrington, P., Support vector machine classification trees based on fuzzy entropy of classification. *Analytica chimica acta* **2017**, 954, 14-21.
23. Selander, E.; Heuschele, J.; Nylund, G. M.; Pohnert, G.; Pavia, H.; Bjærke, O.; Pender-Healy, L. A.; Tiselius, P.; Kjørboe, T., Solid phase extraction and metabolic profiling of exudates from living copepods. *PeerJ* **2016**, 4, e1529.
24. Cortes, C.; Vapnik, V., Support-vector networks. *Machine learning* **1995**, 20 (3), 273-297.

25. Nasrabadi, N. M., Pattern recognition and machine learning. *Journal of electronic imaging* **2007**, *16* (4), 049901.
26. Xu, Z.; Sun, X.; Harrington, P. d. B., Baseline Correction Method Using an Orthogonal Basis for Gas Chromatography/Mass Spectrometry Data. *Analytical Chemistry* **2011**, *83* (19), 7464-7471.
27. de Boves Harrington, P., Statistical validation of classification and calibration models using bootstrapped Latin partitions. *TrAC Trends in Analytical Chemistry* **2006**, *25* (11), 1112-1124.
28. Harrington, P. d. B., Multiple Versus Single Set Validation of Multivariate Models to Avoid Mistakes. *Critical reviews in analytical chemistry* **2018**, *48* (1), 33-46.
29. Harrington, P. d. B.; Vieira, N. E.; Espinoza, J.; Nien, J. K.; Romero, R.; Yergey, A. L., Analysis of variance–principal component analysis: A soft tool for proteomic discovery. *Analytica Chimica Acta* **2005**, *544* (1), 118-127.

OHIO
UNIVERSITY

Thesis and Dissertation Services